

L5

Triggering action potential and electroporation in excitable cells exposed to electric pulses

Tina Batista Napotnik, Lea Rems

University of Ljubljana, Faculty of Electrical Engineering, Slovenia

Duration of the experiments: 60 min

Max. number of participants: 4

Location: Cell Culture Laboratory 3

Level: Basic

PREREQUISITES

Participants should be familiar with Laboratory safety (S1). No other specific knowledge is required for this laboratory practice.

The aim of this laboratory practice is to observe triggering action potentials and electroporation in genetically engineered tet-on spiking human embryonic kidney (S-HEK) cells with the use of 100 μ s electric pulses.

THEORETICAL BACKGROUND

Electric pulses have been used for triggering action potentials in excitable cells (electrostimulation) already for decades. However, when using higher electric fields, the cells' plasma membrane becomes permeabilized and additional ionic currents through pores/defects occurs. These additional ionic currents affect cell excitability in a complex interplay between excitation and electroporation [1, 2].

To better understand the underlying mechanisms of electrostimulation and electroporation in excitable cells (nerves, muscles, cardiac), in vitro experimental work is of great importance. Genetically engineered S-HEK cells expressing a minimal complement of sodium and potassium channels (NaV1.5 and Kir2.1) needed for excitability are a simple and convenient excitable cell model for studying excitation and electroporation in vitro [3, 4].

The use of a fluorescent potentiometric dye ElectroFluor630 and fluorescence microscopy is an effective way to study responses in transmembrane voltage to electric pulses in excitable cells (action potentials and electroporation). Compared to classical electrophysiological methods such as patch clamp, these optical measurements are much easier and time efficient, as they do not require special technical skills. Also, high-voltage electric pulses do not interfere with the measurements (i.e. image acquisition).

EXPERIMENT

We will monitor the changes in transmembrane voltage in excitable S-HEK cells (ATCC CRL-3479) using the fluorescent potentiometric dye ElectroFluor630 (Potentiometric Probes) under a fluorescence microscope and observe the triggering of action potentials and electroporation with 100 μ s electric pulses of increasing amplitudes.

Protocol: S-HEK cells will be plated to Lab-Tek chambers (Thermo Fisher Scientific) 3 days before the experiment in concentration of 105 cells per well. To prepare the cells for experiment, label the cells with 12 μ M ElectroFluor630 in DMEM culture medium for 20 min in a refrigerator at 4°C. Wash the cells three times with a Tyrode solution (2 mM KCl, 125 mM NaCl, 2 mM CaCl₂, 1 mM MgCl₂, 10 mM HEPES, 30 mM glucose, pH 7.3) and at the end, add 1.2 ml of low potassium Tyrode solution (0.5mM KCl, 126.5 mM NaCl, 2 mM CaCl₂, 1 mM MgCl₂, 10 mM HEPES, 30 mM glucose, pH 7.3) to cells.

Insert two parallel Pt/Ir wire electrodes with a 5 mm distance between them to the bottom of the Lab-Tek chamber. Place the chamber under a fluorescence microscope (Thunder Imager Live Cell system for fluorescence microscopy, Leica Microsystems) and connect the electrodes to the pulse generator. Acquire a set of fluorescence images (635 nm excitation and 700 emission wavelength, $\times 40$ objective) in time-lapse acquisition mode, using sCMOS Leica fast camera and LAS X software: acquire 80 images, one image every 36 ms (around 2.8 s total duration of image acquisition). The first image acquisition represents control without pulse application. Further on, while recording, apply a single electric pulse of 100 μ s and 63 V using a TTL signal from the microscope system that triggers the pulse generator and observe the fluorescence signal from the whole field of view.

Every two minutes, record a time-lapse in the same way as before but apply electric pulses of increasing voltage: 63, 75, 88, 100, 125, 150, 175, and 200 V. The electric field to which the cells are exposed is estimated as the applied voltage-to-electrode-distance ratio ($E \approx 126, 150, 176, 200, 250, 300, 350$, and 400 V/cm, respectively). At lower electric fields, the pulses will trigger action potentials, at higher electric fields, the pulses will cause electroporation which manifests as a prolonged depolarization.

The fluorescence signal can be further analyzed using a Matlab application: the acquired images are thresholded to extract the signal only from the membranes, the fluorescence signal is corrected for fading and the characteristic parameters of the signal are extracted (number of action potentials, amplitude, recovery to the baseline etc.), as described more in detail in [1].




Figure 1. Monitoring triggering excitation and electroporation with 100 μ s electric pulses optically with the use of the fluorescent potentiometric probe ElectroFluor630 in excitable S-HEK cells. A brightfield (A) and fluorescence image of S-HEK cells. (C) Time course of ElectroFluor630 fluorescence change in S-HEK cells after applying 100 μ s electric pulses of different amplitudes (signal from the whole field of view, thresholded and corrected for fading).

REFERENCES:

1. Batista Napotnik T., Kos B., Jarm T., Miklavčič D., O'Connor R.P., Rems L. Genetically engineered HEK cells as a valuable tool for studying electroporation in excitable cells. *Sci Rep*, 14:720, 2024.
2. Pakhomov A.G., Pakhomova O.N. The interplay of excitation and electroporation in nanosecond pulse stimulation. *Bioelectrochemistry*, 136:107598, 2020.
3. McNamara H.M., Zhang H., Werley C.A., Cohen A.E. Optically controlled oscillators in an engineered bioelectric tissue. *Phys Rev X*, 6:031001, 2016.
4. Tian H., Davis H.C., Wong-Campos J.D. et al. Video-based pooled screening yields improved far-red genetically encoded voltage indicators. *Nature Methods*, 20:1082-1094, 2023.

NOTES & RESULTS

